



EUROPEAN UNION

Delegation of the European Union to Bosnia and Herzegovina

### **Corrigendum No. 2**

**Project Title:** Supplies of Vehicles, Monitoring and Laboratory Equipment, ICT Equipment and GIS for Inspectorates and Agriculture and Rural Sectors in BiH

**Publication reference:** EuropeAid/126650/C/SUP/BA

**Tender No:** EC/BiH/08/014

1. The technical specification for **Lot 6, Item 6.4 Noise Monitor** shall read as follows:

<b>6.4</b>	<b>Noise monitor</b> <ul style="list-style-type: none"><li>– Class I</li><li>– Range 30 - 130 dB, 12.5Hz - 20kHz</li><li>– A,B,C,D filter</li><li>– 1/3 octave real time analyzer</li><li>– Resolution 0.1 dB</li><li>– Microphone</li><li>– Memory 2 MB</li><li>– USB</li><li>– Battery power supply</li></ul>	7			
------------	---	---	--	--	--

2. The technical specification for **Lot 6, Item 6.7 Handheld pH meter** shall read as follows:

<b>6.7</b>	<b>Handheld pH meter</b> <ul style="list-style-type: none"><li>– Waterproof</li><li>– Range: 1.0 - 15.00 pH</li><li>– Resolution: 0.01 pH</li><li>– Accuracy: 0.01 pH</li><li>– Operating temperature: 0 - 50°C</li><li>– pH buffer standard solutions</li><li>– Battery charged</li></ul>	6			
------------	--	---	--	--	--

3. The technical specifications for **Lot 8, Item 8.2 Laboratory heating mantle (electro mangle), fast** shall read as follows:

8.2	<b>Laboratory heating mantle (electro mantle), fast</b> <ul style="list-style-type: none"> <li>– 300 W</li> <li>– Thermostatic control</li> <li>– Adequate container for heating of cereal samples, 250 ml</li> <li>– One container holder</li> </ul>	7			
-----	---	---	--	--	--