



EUROPEAN UNION

Delegation of the European Union to Bosnia and Herzegovina

Sarajevo, 18<sup>th</sup> March 2010

**Corrigendum No 1**

**Supply of equipment for Microbiology Laboratories in  
Bosnia-and-Herzegovina**

**Tender No: EC/BiH/09/040; EuropeAid/128732/C/SUP/BA**

Following changes in **Annex II+Annex III Technical Specifications and Technical offer** have been made, as listed below:

▪ **Lot 1, Item 1.01 Air handling chamber**

Specifications required for Item 1.01 are completed as follows:

**Hot battery is needed**

**Cold battery is needed**

**Type of batteries needed: water + glycol**

▪ **Lot 2, Item 2.03 Water Jacketed CO2 & O2 Incubator**

Please note that Item 2.03, bullet point saying "Built in CO2/O2 sensors" is corrected and **reads now:**  
**"Built in CO2 sensor".**

▪ **Lot2, Item 2.08 Standard illumination microscope with imaging**

Item 2.08, bullet point saying "Multiple Condensers: at least Abbe, Achromat, phase contrast, polarisation contracts, darkfield dry, darkfield oil" is corrected and **reads now:**

**"Multiple Condensers: at least Abbe, Achromat, darkfield dry, darkfield oil ".**

▪ **Lot 2, Item 2.09 Homogenizer**

Item 2.09, bullet point saying "For samples up to 10 L" is corrected and **reads now:**  
**"For samples up to 8 L".**

▪ **Lot 2, Item 2.11 Deionizer**

Item 2.11, bullet point saying "Treated water to be with TOC<30ppb, Silicates>99.9 %, resistivity<5Mohm" is corrected and **reads now: "Treated water to be with TOC<30ppb, Silicates>99.9 %, resistivity>5Mohm."**

▪ **Lot 2, Lot 2, Item 2.22 : Automated Bacteria Cell Counter**

Specifications required for Item 2.22 are completed as follows:

**Solid Samples approx 9000 pcs/g**