



EUROPEAN UNION

Delegation of the European Commission to Bosnia and Herzegovina

Sarajevo, August 2009

## **CORRIGENDUM No. 2**

### **Project title: General Equipment for VET Schools/centres**

Publication ref: EuropeAid/127150/C/SUP/BA

Tender No: EC/BiH/08/020

#### **1. ANNEX II + III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER**

Lot 6, items 6.23 and 6.39 shall read:

<b>ID No.</b>	<b>Specifications</b>	<b>Qty</b>
<b>6.23</b>	<b>UV/VIS Spectrophotometer</b> Wavelength range: 325 to 1000 nm Bandwidths: 5nm Wavelength resolution: 0.1 nm <b>Stray light: 0.05% at 340 nm</b> Photometric range: <ul style="list-style-type: none"><li>• -0.3 to 2.0 A</li><li>• <b>0 to 199.9%T</b></li><li>• 0 to 9999C</li></ul> Photometric precision: 1% or 0.005A Automatic self test and calibration Lamp selection: automatic, coupled to wavelength PC connection	<b>3</b>
<b>6.39</b>	<b>Polar meter</b> Monochromatic light source: 589 nm Sodium vapour lamp <b>Measuring range: <math>\pm 180^\circ</math></b> Precision: 0.05 % Scale division: $1^\circ$ Polar metric glass tubes: 100nm and 200nm	<b>1</b>

Marc Vanbrabant  
Head of Contracts and Finance